



### Test Conditions

Monitor both the humidity and temperature conditions in a loaded HXD2-1106-01 for a period of 22 hours.

### Test Results

The humidity level in the cabinet was maintained at an average of 0.58% RH. The cabinet was accessed twice, once at 11:47 and again at 13:20, and recovered back to > 1% RH within a 5 min period.

The temperature was set at 42 Deg C and remained between 41.93 - 42.48.

	RH ( +/- 0.5% Tol.)	Temp (+/- 5 C)
Set Point	1% RH	40.0 C
Average	0.48 % RH	41.93 C
Minimum*	0.35 % RH	41.1 C
Maximum*	3.74 % RH	42.48 C

\* After cabinet stabilization at 7:04 AM

\* Vaisala DM70